



ANALIZA BIAŁEK



100002759

Laboratory of Analysis

Client: Marika Mej
Order number: 100002759
Name: NASSA LABS DROSTANOLONE ENANTHATE
Number of samples: 1 vial

Batch number: -----
Recive date: 2022-09-09
Start date: 2022-09-16
End date: 2022-09-20

Sample identification

Substance	Method of taking a sample	Form
Drostanolone Enanthate	Sample provided by customer	Light yellow liquid in a glass vial

Results

Substance	Sample name	Content	Purity
Drostanolone Enanthate	NASSA LABS DROSTANOLONE ENANTHATE 200MG/ML	199,04mg/ml	-----

Research comment

The test sample was applied in the amount of 10 μ l to the BioBasic 8 Thermo Scientific column. During the course of the analysis with the method a linear gradient in the solutions of H₂O and Acetonitrile with the addition of 0.1% trifluoroacetic acid, a peak of the sought substance appeared. Its concentration, when compared to the reference chromatogram, is 199,04mg/ml. Height the measurement error is + - 1 mg. The actual state of the sample sent before the test is presented on the photo below

Photo of the test sample



Password

354GSDAM


Analiza Białek sp. z o.o.
KRS: 0000966168
NIP: 8992922970
REGON: 521751875

Siedziba i Laboratorium:
ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Badanie przeprowadził:
Specjalista ds. analiz i badań
mgr Arkadiusz Zajęc



Chromatogram



Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0min	60%	40%
2min	60%	40%
27min	5%	95%
30min	5%	95%

Password

354GSDAM